

Instruction for Use

Product: Lap Bag with Auto-Retrieval System

The Following information should be read before using this device.

Indications:

The Lap Bag with Auto-Retrieval system has applications in minimally invasive abdominal procedures for the safe, convenient removal of tissue specimens such as the appendix, gallbladder, gallstones, lymph nodes, an ectopic pregnancy, ovaries, small sections of bowel and other tissue structures.



Contraindications:

The instrument is not intended for use when minimally invasive techniques are contraindicated.

Device Description:

The Lap Bag with Auto-Retrieval system is an elastic pouch that minimizes spillage and intraoperative contamination by isolating and containing tissue specimens. The Lap Bag with Auto-Retrieval system is designed for introduction and use through 10/11 and 12 mm cannula.

REF#	Bag Size	Volume capacity
ALB-AV30	3 x 6 inch	170 ml
ALB-AV50	5 x 7 inch	260 ml
ALB-CUD0	4 x 4 inch	150 ml
ALB-CUE0	5 x 5 inch	300 ml
ALB-CUF0	6 x 6 inch	800 ml
ALB-CUG0	8 x 8 inch	1250 ml

*Note: User shall evaluate the condition and choose the appropriate device.

Direction for Use:

1. Using sterile technique, remove the instrument from the packaging. To avoid damage, do not flip the instrument into the non-sterile field.
2. Take warning label away.
3. Insert the Lap Bag with Auto-Retrieval system into the abdomen and pushing through an appropriately sized trocar cannula. The bag will open automatically.
4. Place the specimen into the pouch.
5. Once the specimen is completely within the confines of the pouch, straight up pull the rear handle away from the flexible cannula (this step will keep the bag closed and take apart from the rigid handle).
6. Take the flexible cannula away through Trocar cannula.

Removing the Pouch and Trocar Cannula.

- The suture with O-ring will keep in outside of abdomen.
- The pouch with specimen can be removed through a 10 mm trocar site, for specimens approximating this diameter.
- Pull back on the O-ring of the suture until the bag complete closed, neck of the pouch reaches the body surface. Visually inspect to see if pouch reaches the body surface, if air is trapped in the pouch, use an atraumatic instrument to enlarge the neck of the pouch to allow air to escape.

- If pouch does not pull through easily, enlarge the incision to accommodate easy removal of the pouch, taking care not to puncture the pouch.

Caution: Do not attempt to pull the pouch device through the trocar cannula when the tissue sample is larger than the cannula diameter. Please mince the tissue to avoid the formation of a pouch with large tissue fragments stuck at the edge of the incision port.

Precautions/Warnings:

- This device is designed for single use only and is not to be re-used or re-sterilized. Do not use if package is opened or damaged.
- Contents sterile unless enclosed package has been opened or damaged. Store at 15°C~25°C.
- Endoscopic procedures should be performed only by physicians with adequate training and knowledge of these procedures. In addition, medical literature should be consulted regarding techniques, hazards, contraindications and complications prior to the performance of these procedures.
- Before endoscopic instruments and accessories from different manufactures are utilized together in a procedure, verify compatibility.
- This device is not intended for use with any tissue that will not fit within the confines of the pouch to allow complete closure.
- Do not remove, or attempt to remove the specimen pouch through the trocar cannula.
- Not for use with Power morcellators! The Lap Bag is not intended for use with sharp surgical instruments. Care should be taken to avoid damaging, puncturing or tearing the Lap Bag.
- Beware of the remaining heat on the surgical instrument might damage the pouch.
- Dispose used product following local state or federal guidelines.



Caution

STERILE EO

Sterilized using Ethylene Oxide



Consult instructions for use



Do not use if package is damaged



Batch code



Do not reuse



Keep away from sunlight



Date of manufacture



Keep dry



Catalogue number



Manufacturer



Temperature limitation: 15°C~25°C



Do not re-sterilize



Use by



Single sterile barrier system with protective packaging outside



Medical device



Single sterile barrier system

If further information is required, please contact:



Advanced Medical Design Co., Ltd.
4F~5F, No29, Wuquan 5th Rd., Wugu Dist.,
New Taipei City, 24888 Taiwan R.O.C.
TEL: 886-2-22902627
FAX: 886-2-22988501
<https://www.amdintl.com.tw/>
Email: sales@amdintl.com.tw